

Validation of a Fully Automatic Photoplethysmographic Device for Toe Blood Pressure Measurement

A. Pérez-Martin ^{a,*}, G. Meyer ^a, C. Demattei ^b, G. Böge ^{a,c}, J.-P. Laroche ^c, I. Quéré ^c, M. Dazat ^a

^a EA2992 Dynamics of Cardiovascular Incoherencies Laboratory, Montpellier I University, Nîmes and the Department of Vascular Medicine, Nîmes University Hospital, Nîmes, France

^b Department of Biostatistics, Clinical Epidemiology, Public Health and Medical Information, Nîmes University Hospital, Nîmes, France

^c Vascular Medicine Unit, Montpellier University Hospital, Montpellier, France

Submitted 11 May 2010; accepted 13 June 2010
Available online 8 July 2010

KEYWORDS

Toe blood pressure;
Photoplethysmography;
Automation;
Validation

Abstract Objectives: This study was designed to assess the accuracy and reliability of a new, portable, fully automated photoplethysmography (PPG) device for toe blood pressure (TBP) measurement.

Design: Within-subject comparison with conventional laser Doppler (LD) measurement.

Materials and methods: Four TBP measurements were performed on both lower limbs, alternatively with LD and PPG in 200 patients recruited at the Nîmes University Hospital. Reproducibility was assessed by the intraclass correlation coefficient (ICC). The concordance between the two methods was evaluated by Lin's concordance correlation coefficient (CCC), in the whole population as well as in comorbidity subgroups. A potential bias was investigated with the Bland and Altman method.

Results: The ICC was 0.887 (95% confidence interval (CI) 0.852–0.913) and 0.893 (0.860–0.918) on the right side ($n = 193$), 0.905 (0.875–0.928) and 0.898 (0.866–0.922) on the left side ($n = 188$) for PPG and LD measurements, respectively. The CCC was 0.913 (0.885–0.934) on the right side and 0.915 (0.888–0.937) on the left side, and remained >0.8 regardless of co-morbidities.

Conclusions: This new, fully automatic, photoplethysmographic device yielded reliable TBP measurements and showed good agreement with the reference LD system over a wide range of values.

© 2010 European Society for Vascular Surgery. Published by Elsevier Ltd. All rights reserved.

* Corresponding author at: Service d'Exploration et Médecine Vasculaire, CHU de Nîmes, Place du Pr R Debré, 30029 Nîmes Cedex 9, France. Tel.: +33 (0) 466683313; fax: +33 (0) 466684222.

E-mail address: antonia.perez.martin@chu-nimes.fr (A. Pérez-Martin).

Peripheral arterial disease (PAD) is a common consequence of systemic atherosclerosis, but may remain clinically silent, especially in sedentary subjects or in diabetic patients with peripheral neuropathy. Asymptomatic forms are thus believed to be more common than symptomatic forms, but both are associated with an increased risk of cardiovascular mortality and morbidity, in addition to local trophic complications.^{1–3} Therefore, screening for PAD appears to be of importance and requires an accurate, quick and cost-effective diagnostic approach.

The ankle brachial pressure index (ABI) is a sensitive and specific method⁴ not only for diagnosing but also for assessing the severity of PAD,⁵ and is largely recommended.⁶

However, ABI measurement requires specific tools (blood pressure cuffs and handheld 5–10-MHz continuous-wave Doppler probe), is operator dependent, time consuming⁷ and remains underused, especially in outpatients. Moreover, its reliability is limited or compromised when medial calcinosis reduces peripheral artery compressibility, especially in patients with diabetes, chronic renal failure or advanced age, in whom the risk of PAD is greater.^{8,9}

Measuring systolic blood pressure at the toe (TBP) and calculating the toe-brachial index (TBI) has been proposed to overcome this limitation because small distal arteries are less prone to medial calcinosis.^{10–13} Although some authors consider the TBI useful only in patients with elevated ABI (i.e., >1.3),¹⁰ others emphasise the risk of falsely normal ABI in patients with moderate incompressibility of ankle arteries.¹¹

However, TBP measurement also requires specific equipment, including small pneumatic cuffs and plethysmography or laser Doppler (LD) devices, and its validity depends on its use by skilled and meticulous operators. Several specialised systems have been developed, the more recently validated using laser Doppler.¹⁴ So far, the use of such devices on a large-scale has been hampered by their complexity and/or cost, so that TBP is often regarded as unsuitable for the outpatient clinic.

This is why we developed a new miniaturised and fully automated TBP measurement device using photo-plethysmography (PPG) flow detection, with automatic pulp blood draining as well as occlusion cuff inflation and deflation. We assessed the usability, accuracy and reliability of this device in a large sample of patients with and without cardiovascular risk factors or expected PAD, using a within-subject comparative design with conventional, LD laboratory equipment.

Materials and Methods

Subjects

A 200-subject study population sample was recruited at the Nîmes University Hospital over a 4-month period from in- and outpatients aged 18 years or over: a total of 50 consecutive diabetic patients admitted to the Department of Diabetes and Metabolic Medicine, 50 patients from the Department of Internal Medicine and 100 patients referred to the Department of Vascular Medicine. Exclusion criteria were marked limb tremor or skin ulceration on toe

measurement sites. All participants gave written informed consent. The study protocol was approved by the local Ethics Committee and conducted in accordance with the Declaration of Helsinki.

Experimental procedure

Room temperature was maintained at 24–25 °C. Individuals were rested in a supine position and acclimatised for at least 10 min. Information on patient demographics and comorbidities was collected with a questionnaire and medical record analysis, especially as regards diabetes, hypertension, PAD, distal ulcer and smoking habits.

Brachial blood pressure

Systolic and diastolic blood pressures were measured at rest in supine position, using an automated system (Dinamap DPC-320N, GE Medical Systems, Milwaukee, WI, USA).

Toe pressure measurement

All measurements were performed on the first toe of both lower limbs, if available, or on the second toe, if not. Prior to blood pressure measurement, one sequence of inflation–deflation of the pneumatic cuffs was completed but not recorded to avoid or limit post-ischaemic effects. Toe systolic pressure assessments were repeated for a total of four measurements, twice with the new PPG device – SysToe[®] (Atys Medical, Soucieu-en-Jarrest, France, Patent # 08 58291, 5 December 2009), and twice with the reference LD system (PF 5010 Doppler Laser monitor, Perimed, Stockholm, Sweden), the techniques being used alternatively in a randomised order, one technique followed by the other one (e.g., LD, PPG, LD, PPG or PPG, LD, PPG, LD).

TBP measurement with LD was conducted according to the manufacturer's instructions. After wrapping the appropriately sized pneumatic cuff around the first phalanx of the toe, the laser probe was placed distal to the cuff on the toe pulp. Pulp blood draining was performed manually by gentle compression immediately before inflating the occlusion cuff to 250 mmHg. The cuff was then automatically and linearly deflated. Systolic blood pressure was read automatically by software when the LD probe detected the return of blood perfusion (Fig. 1).

For blood pressure measurement with the PPG device, an appropriately sized pneumatic cuff (occlusion cuff) was wrapped around the first phalanx of the toe. The PPG sensor was placed on the toe pulp and maintained with a second pneumatic cuff (blood draining cuff) delicately wrapped so that it fitted the toe without applying any pressure when deflated. Turning on the device launched an automatic sequence starting with inflation of the blood draining cuff, then inflation of the occlusion cuff followed by deflation of the blood draining cuff, after which the occlusion cuff was slowly and linearly deflated. The infrared PPG system, working in reflection mode at 880 nm wavelength, with a fixed power of 75 mW, provided penetration depth of about 3 mm in the skin. During deflation, the return of blood perfusion was detected when the direct current (DC) curve showed a continuous upslope indicating a steady increase in toe blood volume (Fig. 1). The pressure measured within the occlusion cuff at the beginning of this upslope was assumed to correspond to local systolic blood pressure.




Figure 1 Print of a typical automatic toe blood pressure measurement with laser Doppler (LD) and photoplethysmography (PPG) in a patient with peripheral arterial disease. Time is shown on the horizontal axis. On the LD curve (upper part of the graph), the figures along the horizontal axis correspond to the instantaneous pressure in the occlusion cuff, whereas the curve shows the LD signal. On PPG curve (lower part of the graph), the blue curve shows the pressure in the occlusion cuff, read in mm Hg on the left scale, whereas the red curve shows the PPG signal in % of its maximum amplitude change during the measurement sequence.

Statistical analysis

For each measurement device and each lower limb, reproducibility was assessed by the intraclass correlation coefficient (ICC) between the two repeated pressure measurements. This reliability index was computed as described by Carrasco *et al.*,¹⁵ the individual effect in the model being considered as a random effect. The measurement model was thus $y_{ij} = \mu + u_i + \varepsilon_{ij}$ in which $\varepsilon_{ij} \sim N(0, \sigma^2)$ and $u_i \sim N(0, \tau^2)$ are independent and y_{ij} denotes the measurement for individual i at occasion j . This model decomposes the total variance between error variance (σ^2) and inter-individual variance (τ^2). Then, $ICC = \tau^2 / (\tau^2 + \sigma^2)$ is the proportion of the total variance that is due to subjects. The reproducibility of a method of measurement is judged acceptable when the ICC is higher than 0.8.

The concordance between the two methods of measurement was assessed by the concordance correlation coefficient (CCC) of Lin.¹⁶ The concordance computation was based on the average of the two measurements done by each device on each lower limb if the devices were separately judged reproducible, and on the first and second measurement separately, otherwise.

The Bland and Altman method was used to search for a potential bias between the devices.

Reproducibility of measurements and concordance of devices were assessed on the whole sample, by department, and on subsamples defined by co-morbidities. Reproducibility and concordance were also assessed for TBI.

Each of the two lower limbs was treated separately to avoid an additional source of variability.

All coefficient estimates are given with their 95% confidence interval (CI). Data management and statistical

analysis were performed under R software version 2.9.2 (R Development Core Team (2009), R Foundation for Statistical Computing, Vienna, Austria.)

Results

LD measurement could not be obtained in two subjects because of tremor artefacts. PPG measurements were possible in all 200 subjects. TBP measurement could not be performed by either method in five right lower limbs and in 10 left lower limbs where toes or legs had been amputated. Therefore, 193 right lower limbs and 188 left lower limbs, in 114 males and 84 females, were available for comparison. There were 109 (55%) diabetic patients, 51 (25.6%) patients with identified PAD, 82 (41.4%) patients with arterial hypertension, 23 (11.6%) patients with distal ulcers and 58 (29.3%) smokers (Table 1).

Although the average time for measurements was similar between the two systems, the software procedure was longer with the LD device (software initialisation and patients' identification) than with PPG system. Typically; a series of measurements lasts about 5 min with the PPG versus about 10 min with the LD system. For the present study, the entire procedure (i.e., two measurements with each device) took about 15 min to complete.

The ICC was 0.887 (95% CI 0.852–0.913) and 0.893 (0.860–0.918) on the right side, 0.905 (0.875–0.928) and 0.898 (0.866–0.922) on the left side, respectively, for PPG and LD measurements.

The CCC of systolic pressure and TBI for all subjects and for each subgroup are reported in Table 2. Concordance was fair (CCC >0.8), or good (>0.9) in most subgroups. The analysis was not done for patients with ulcers, because the number of subjects in this subgroup was too small. The concordance was not altered by systolic pressure level and remained above 0.8, regardless of brachial pressure values.

Fig. 2 shows LD versus PPG measurements for each lower limb. The 45° line through the origin represents the perfect situation of concordance. The best linear regression model explaining LD measurement values as a function of PPG values was the model without a constant. The estimated coefficients (0.922 for the right side and 0.939 for the left side) were significantly different from 0 ($p < 0.001$) and close to 1. The coefficient of determination (R^2) for the right and left models were 0.984 and 0.983, respectively.

PPG measurements yielded mean pressure values 7 mmHg higher than LD. However, the Bland and Altman analysis showed that the difference between the two methods did not depend on toe pressure values (Fig. 3).

Discussion

This study demonstrated that the fully automatic portable PPG device that we developed yielded reliable TBP measurements in patients, regardless of their co-morbidities, and showed good agreement with the reference LD system over a wide range of values.

PPG measurement yielded, on average, slightly higher TBP values than LD, although previous studies have reported the opposite.¹⁷ LD is regarded as highly sensitive and may be able to detect flow in patients with very low TBP in

Table 1 Subjects characteristics.

	Total <i>n</i> = 198	DMMD <i>n</i> = 50	IMD <i>n</i> = 50	VMD <i>n</i> = 98
Men/women ratio	114/84	60/38	23/27	30/20
Age (years) [mean ± SD]	63.78 ± 14.90	65.15 ± 13.59	57.40 ± 14.96	67.55 ± 15.63
PAD [<i>n</i> (%)]	51 (25.6%)	7 (14%)	7 (14%)	37 (37.8%)
Diabetes [<i>n</i> (%)]	109 (55%)	50 (100%)	11 (22%)	48 (49%)
Insulin treated [<i>n</i> (% of diabetics)]	64 (58.7%)	29 (58%)	7 (63.6%)	14 (21.1%)
Hypertension [<i>n</i> (%)]	82 (41.4%)	27 (54%)	9 (18%)	46 (46.9%)
Distal ulcer [<i>n</i> (%)]	23 (11.6%)	1 (2%)	15 (30%)	7 (7.1%)
Smokers [<i>n</i> (%)]	58 (29.3%)	17 (34%)	8 (16%)	46 (46.9%)

DMMD: Diabetes and Metabolic Medicine Department, IMD: Internal Medicine Department, VMD: Vascular Medicine Department, PAD: Peripheral Arterial Disease.

whom PPG would detect no flow.¹⁷ Nevertheless, there was no case in our series where LD could detect the return of blood perfusion while PPG could not. In addition, all patients with values below the 30 mmHg-threshold defining ischaemia were detected by both devices. Moreover, the Bland and Altman analysis showed that the difference between the two devices was not greater for low than for normal or high TBP values. Finally, in the PAD subgroup, the CCC remained satisfactory: 0.869 (0.783–0.922) and 0.911 (0.844–0.950) on the right and left sides, respectively. Therefore, this small difference may simply reflect the microcirculation level reached by each technique, with LD recording flow in smaller vessels in which blood pressure is lower. From a clinical point of view, this difference seems irrelevant and PPG appears to be discriminating enough to detect high-risk subjects.

On the other hand, previous studies¹⁸ reported large variations between PPG and LD, whereas our results showed a much closer agreement between these two techniques. This is probably because of the improved sensitivity and reliability of both state-of-the-art sensors we used in our study.

PPG signals may be influenced by many physiological inputs, including respiration, thermoregulation and sympathetic nervous system activity,¹⁹ which may affect

the pulsatile plethysmography curve recorded in the alternating current (AC) mode, but not the DC curve that was used by our system to detect the steady upslope resulting from blood flow return in the toe pulp.

PAD, even if asymptomatic, is a prognostic marker for total mortality and cardiovascular events.^{2,18} Therefore, its early detection could allow improving risk factor control and reducing the rate of cardiovascular events, thus ensuring a healthier outcome. In high-risk patients, especially in diabetics, close follow-up is mandatory. Therefore, improving the screening, diagnosis and follow-up of PAD represents a challenge requiring accurate, reliable, simple, accessible and non-invasive tools. Doppler-measured ABI, although validated and recommended,⁶ has significant limitations and drawbacks, including the required training and costs.²⁰ Some authors have tested automatic procedures using toe-applied oscillometry,²¹ oximetry^{12,22} or PPG for measuring ankle blood pressure.²³ However, none of these techniques has so far been proven suitable for large-scale use. In addition, the validity of ABI measurement is compromised in patients with medial calcinosis, in whom TBP measurement is more appropriate,⁸ especially when ABI is ≥ 1.3 .¹⁰

TBI assessment requires specific, delicate, more or less cumbersome and costly equipment, which has so far

Table 2 LD and PPG measurements of toe pressure and toe-brachial systolic pressure index: concordance correlation coefficient (CCC) and 95% confidence interval (95%CI) in the whole population sample and in comorbidity subgroups.

	Right lower limb CCC (95%CI)	Left lower limb CCC (95%CI)
Toe blood pressure		
Total (<i>n</i> = 198)	0.913 (0.885–0.934)	0.915 (0.888–0.937)
DMMD (<i>n</i> = 50)	0.913 (0.851–0.950)	0.9165 (0.854–0.953)
IMD (<i>n</i> = 50)	0.948 (0.907–0.971)	0.919 (0.854–0.956)
VMD (<i>n</i> = 98)	0.8969 (0.848–0.931)	0.9091 (0.864–0.940)
PAD (<i>n</i> = 51)	0.869 (0.783–0.922)	0.911 (0.844–0.950)
Diabetes (<i>n</i> = 109)	0.9059 (0.864–0.936)	0.9061 (0.862–0.936)
Hypertension (<i>n</i> = 82)	0.912 (0.865–0.943)	0.913 (0.865–0.945)
Smokers (<i>n</i> = 58)	0.920 (0.864–0.953)	0.918 (0.861–0.953)
Toe-brachial systolic pressure index		
Total (<i>n</i> = 198)	0.889 (0.853–0.916)	0.891 (0.856–0.918)

DMMD: Diabetes and Metabolic Medicine Department, IMD: Internal Medicine Department, VMD: Vascular Medicine Department, PAD: Peripheral Arterial Disease.




Figure 2 Plot of laser Doppler (LD) toe pressure measurement versus photoplethysmography (PPG) toe pressure measurements (A) for the right lower limb and (B) for the left lower limb. The full line is the 45° line through the origin which represents the perfect situation of concordance. The best linear regression model explaining LD measurement values as a function of PPG values was the model without constant. The equation of the regression line (dotted line) was $LD = 0.922 PPG$ for the right lower limb and $LD = 0.939 PPG$ for the left lower limb.




Figure 3 Bland–Altman plot showing the difference between photoplethysmography (PPG) and laser Doppler (LD) toe blood pressure measurement at the right (A) and the left (B) lower limb in dependency of the mean value of PPG and LD in each measurement.

precluded its use in out-clinics, outside vascular laboratories. Several devices are available, most of them using either PPG or LD sensors. During the past few years, some authors have developed automated monitors for TBP assessment.²⁴ Among them, the PresTo[®], first validated in 2002, has been more recently tested in 16 diabetic patients with ischaemic leg by comparing the automatic software detection of blood perfusion return to the visual assessment of LD signals.²⁵ In the present study, we compared values provided by automatic software detection of blood perfusion return on PPG and on LD signal, and obtained quite satisfactory results.

In addition, the small PPG device we used is portable, battery powered, with fully automated reading of blood pressure, and also of pulp blood draining and occlusion cuff inflation and deflation, which are either not implemented or manual on the other systems.

We tested the accuracy of our device over a large range of pressure values, in patients with or without PAD and comorbidities or cardiovascular risk factors. In all groups and subgroups, the reliability suited the clinical needs, and the reproducibility was satisfactory.

Although this device requires little formal training, its reliability implies the adherence to strict methodological recommendations such as warm room temperature and supine rest. The cuff width also influences the toe pressure value and needs to be taken into account in patients with a very large first toe, or when measurements have to be performed on another toe because of first toe amputation or ulceration.²⁶ In addition, the PPG sensor must be precisely placed with full and stable contact with the skin. The use of two-sided adhesive tape is useful.

In conclusion, fully automated PPG measurement of TBP with a small portable device is reliable and accurate. This technique seems perfectly suitable for large-scale screening, evaluation, and follow-up of PAD in at-risk patients, particularly diabetics.

Conflict of Interest

None of the authors of the present study and none of their relatives received, is receiving, or will receive any direct or indirect reward for this work besides her or his normal salary from the university and/or hospital centre. None of the co-authors of the present study and none of their relatives will receive any benefit or advantage from the commercialisation and sales of the SysToe[®] device.

Acknowledgements

The development of the SysToe[®] device, and the present clinical study, were supported by a grant from the ANR (French National Research Agency) # ANR05RNTS01701. We thank Dr. Carey Suehs for her help with English editing.

References

- 1 Heald CL, Fowkes FG, Murray GD, Price JF. Risk of mortality and cardiovascular disease associated with the ankle-brachial index: systematic review. *Atherosclerosis* 2006;**189**:61–9.
- 2 Criqui MH, Langer RD, Fronek A, Feigelson HS, Klauber MR, McCann TJ, et al. Mortality over a period of 10 years in patients with peripheral arterial disease. *N Engl J Med* 1992;**326**:381–6.
- 3 Ramos R, Quesada M, Solanas P, Subirana I, Sala J, Vila J, et al. Prevalence of symptomatic and asymptomatic peripheral arterial disease and the value of the ankle-brachial index to stratify cardiovascular risk. *Eur J Vasc Endovasc Surg* 2009;**38**:305–11.
- 4 Yao ST, Hobbs JT, Irvine WT. Ankle systolic pressure measurements in arterial disease affecting the lower extremities. *Br J Surg* 1969;**56**:676–9.
- 5 Fowkes FG, Murray GD, Butcher I, Heald CL, Lee RJ, Chambless LE, et al. Ankle brachial index combined with Framingham risk score to predict cardiovascular events and mortality: a meta-analysis. *JAMA* 2008;**300**:197–208.
- 6 Hirsch AT, Haskal ZJ, Hertzner NR, Bakal CW, Creager MA, Halperin JL, et al. ACC/AHA guidelines for the management of patients with peripheral arterial disease (lower extremity, renal, mesenteric, and abdominal aortic): a collaborative report from the American Association for Vascular Surgery/Society for Vascular Surgery, Society for Cardiovascular Angiography and Interventions, Society for Vascular Medicine and Biology, Society of Interventional Radiology, and the ACC/AHA Task Force on Practice Guidelines (writing committee to develop guidelines for the management of patients with peripheral arterial disease)—summary of recommendations. *J Vasc Interv Radiol* 2006;**17**:1383–97.
- 7 Nicolai SP, Kruidenier LM, Rouwet EV, Bartelink ML, Prins MH, Teijink JA. Ankle brachial index measurement in primary care: are we doing it right? *Br J Gen Pract* 2009;**59**:422–7.
- 8 Williams DT, Harding KG, Price P. An evaluation of the efficacy of methods used in screening for lower-limb arterial disease in diabetes. *Diabetes Care* 2005;**28**:2206–10.
- 9 Aboyans V, Ho E, Denenberg JO, Ho LA, Natarajan L, Criqui MH. The association between elevated ankle systolic pressures and peripheral occlusive arterial disease in diabetic and nondiabetic subjects. *J Vasc Surg* 2008;**48**:1197–203.
- 10 Brooks B, Dean R, Patel S, Wu B, Molyneaux L, Yue DK. TBI or not TBI: that is the question. Is it better to measure toe pressure than ankle pressure in diabetic patients? *Diabet Med* 2001;**18**:528–32.
- 11 Suominen V, Uurto I, Saarinen J, Venermo M, Salenius J. PAD as a risk factor for mortality Among patients with elevated ABI—a clinical study. *Eur J Vasc Endovasc Surg* 2010;**39**(3):316–22.
- 12 Parameswaran GI, Brand K, Dolan J. Pulse oximetry as a potential screening tool for lower extremity arterial disease in asymptomatic patients with diabetes mellitus. *Arch Intern Med* 2005;**165**:442–6.
- 13 Sahli D, Eliasson B, Svensson M, Blohme G, Eliasson M, Samuelsson P, et al. Assessment of toe blood pressure is an effective screening method to identify diabetes patients with lower extremity arterial disease. *Angiology* 2004;**55**:641–51.
- 14 de Graaff JC, Ubbink DT, Legemate DA, de Haan RJ, Jacobs MJ. The usefulness of a laser Doppler in the measurement of toe blood pressures. *J Vasc Surg* 2000;**32**:1172–9.
- 15 Carrasco JL, Jover L. Estimating the generalized concordance correlation coefficient through variance components. *Biometrics* 2003;**59**:849–58.
- 16 Lin LI. A concordance correlation coefficient to evaluate reproducibility. *Biometrics* 1989;**45**:255–68.
- 17 Ubbink DT. Toe blood pressure measurements in patients suspected of leg ischaemia: a new laser Doppler device compared with photoplethysmography. *Eur J Vasc Endovasc Surg* 2004;**27**:629–34.
- 18 Newman AB, Shemanski L, Manolio TA, Cushman M, Mittelmark M, Polak JF, et al. Ankle-arm index as a predictor of cardiovascular disease and mortality in the cardiovascular health study. The Cardiovascular Health Study Group. *Arterioscler Thromb Vasc Biol* 1999;**19**:538–45.
- 19 Allen J. Photoplethysmography and its application in clinical physiological measurement. *Physiol Meas* 2007;**28**(3):R1–39.
- 20 Mohler 3rd ER, Treat-Jacobson D, Reilly MP, Cunningham KE, Miani M, Criqui MH, et al. Utility and barriers to performance of the ankle-brachial index in primary care practice. *Vasc Med* 2004;**9**:253–60.
- 21 Adiseshiah M, Cross FW, Belsham PA. Ankle blood pressure measured by automatic oscillometry: a comparison with doppler pressure measurements. *Ann R Coll Surg Engl* 1987;**69**:271–3.
- 22 Johansson KE, Marklund BR, Fowelin JH. Evaluation of a new screening method for detecting peripheral arterial disease in a primary health care population of patients with diabetes mellitus. *Diabet Med* 2002;**19**:307–10.
- 23 Jonsson B, Laurent C, Eneling M, Skau T, Lindberg LG. Automatic ankle pressure measurements using PPG in ankle-brachial pressure index determination. *Eur J Vasc Endovasc Surg* 2005;**30**:395–401.
- 24 Wahlberg E, Gush R. A new automated toe blood pressure monitor for assessment of limb ischemia. *Eur J Vasc Endovasc Surg* 2002;**24**:304–8.
- 25 Pahlsson HI, Lund K, Jorreskog G, Gush R, Wahlberg E. The validity and reliability of automated and manually measured toe blood pressure in ischemic legs of diabetic patients. *Eur J Vasc Endovasc Surg* 2008;**36**:576–81.
- 26 Pahlsson HI, Jorreskog G, Wahlberg E. The cuff width influences the toe blood pressure value. *Vasa* 2004;**33**:215–8.